



## DECLARATION OF CONFORMITY Seametrics Inc.

Certifies that the following rate and total indicator products conform with the protection requirements of Council Directive 2004/108/EC on the approximation of the laws of the member states relating to electromagnetic compatibility:

**AG3000, AG3000P, AG3000R, iMAG4700, iMAG4700P, iMAG4700R**

### EMC standards applied:

EN61326-1:2013 Electrical Equipment for Measurement, Control and Laboratory EMC Requirements; General Requirements (Industrial)

EN61326-2-3:2013 Electrical Equipment for Measurement, Control and Laboratory EMC Requirements; Particular Requirements, Test Configurations, operational conditions and performance criteria for transducers with integrated or remote signal conditioning

### Test Methods:

#### *Radiated and Conducted Emissions:*

CISPR11:2015

#### *Immunity (EN 61326-1:2013) Test Methods:*

IEC61000-4-2:2008 ESD

IEC61000-4-3:2010 Radiated Immunity

IEC61000-4-4:2012 EFT

IEC61000-4-5:2014 Surge

IEC61000-4-6:2013 Conducted Immunity

IEC61000-4-8:2009 Magnetic Field Immunity

IEC61000-4-11:2004 Voltage Interruptions



Issued: February 2016

Signed: 

Fred Sutton Engineering Manager