

DECLARATION OF CONFORMITY

Seametrics Inc.

Certifies that the following magmeter products conform with the protection requirements of Council Directive 2014/30/EU and the RoHS requirements of Council Directive (EU) 2015/863:

AG3000, AG3000P, AG3000R, IMAG4700, IMAG4700P, IMAG4700R

EMC standards applied:

EN61326-1:2013 Electrical Equipment for Measurement, Control and Laboratory EMC Requirements; General Requirements (Industrial)

EN61326-2-3:2013 Electrical Equipment for Measurement, Control and Laboratory EMC Requirements; Particular Requirements, Test Configurations, operational conditions and performance criteria for transducers with integrated or remote signal conditioning

Test Methods:

Radiated and Conducted Emissions:

CISPR11:2015

Immunity (EN 61326-1:2013) Test Methods:

IEC61000-4-2:2008 ESD

IEC61000-4-3:2010 Radiated Immunity

IEC61000-4-4:2012 EFT


IEC61000-4-5:2014 Surge

IEC61000-4-6:2013 Conducted Immunity

IEC61000-4-8:2009 Magnetic Field Immunity

IEC61000-4-11:2004 Voltage Interruptions

Issued: October 2019

Signed:  _____

Fred Sutton Engineering Manager